

\*\*\* DRAFT - NOT YET FILED \*\*\*

3745-89-05

**Requirements for maintaining Maintenance of a laboratory certification.**

- (A) The requirements for maintaining laboratory certification include each of the following:
- (1) Conformance by the laboratory to the "Ohio EPA Laboratory Manual for the Microbiological Analyses of Public Drinking Water 2014" and the "Ohio EPA Laboratory Manual for the Chemical Analyses of Public Drinking Water 2014,"
  - (2) Documentation that all analysts designated on the laboratory certificates have participated proportionally in drinking water analyses for which they are certified.
  - (3) Documentation that only individuals who have interim authorization or have participated acceptably in a recent on-site survey for a laboratory certification or are listed on a laboratory certificate have performed drinking water analyses and reported results thereof for the purpose of determining compliance with plant control tests or MMO-MUG (SM 9223) tests, and Chapters 3745-81 ~~and~~, 3745-82; and 3745-90, and rules 3745-91-06 and 3745-9-09, and as specified by paragraph (B) of rule 3745-83-01 of the Administrative Code.
  - (4) Documentation that drinking water analyses required for the purpose of determining compliance with plant control tests and Chapters 3745-81 ~~and~~, 3745-82; and 3745-90, and rules 3745-91-06 and 3745-9-09, and as specified by paragraph (B) of rule 3745-83-01 of the Administrative Code, are performed in accordance with the analytical methods which the laboratory has been certified to use.
  - (5) Obtaining written approval from the director prior to making any changes in the analytical methodologies or procedures which the laboratory has been certified to use.
  - (6) Documentation that the laboratory and equipment therein conform to the laboratory plan approved by the director, and that all necessary laboratory equipment is operational.
  - (7) Notification to the director in writing of any change in personnel or equipment which may affect the laboratory's ability to accurately perform drinking water

analyses.

- (8) Submitting revised plans and obtaining approval from the director prior to any proposed laboratory remodeling or relocation. During any remodeling or relocation, the director may refuse to accept drinking water analytical results for the purpose of determining compliance with plant control tests and Chapters 3745-81 ~~and~~, 3745-82, and 3745-90, and rules 3745-91-06 and 3745-9-09, and as specified by paragraph (B) of rule 3745-83-01 of the Administrative Code. Upon completion of said remodeling or relocation the laboratory shall notify the director in writing.
- (9) Documentation that an acceptable quality assurance plan as described in rule 3745-89-03 of the Administrative Code is being followed.
- (10) Maintenance of an acceptable quality assurance plan for drinking water analyses and submittal of revisions to the director prior to implementation of any revisions.
- (11) An acceptable level of performance for the reporting of analytical data as described in rule 3745-89-08 of the Administrative Code, including an absence of any falsification in the reports, as well as reporting analytical results in accordance with the reporting limits established in appendix B to rule 3745-89-03 of the Administrative Code.
- (12) Obtaining acceptable results within the past twelve months in at least one proficiency test for each regulated contaminant that the laboratory has certification.
- (13) Satisfactory performance for analyses included in interlaboratory comparison and verification studies, if such are required by the director.
- (14) Allowing on-site surveys of the laboratory by certification personnel, including unannounced on-site surveys.
- (15) Acceptable performance as described in paragraph (A)(5) of rule 3745-89-03 of the Administrative Code during each on-site survey, including unannounced on-site surveys.
- (16) Correction of deviations noted in survey reports, as set forth in paragraph (A)(6) of rule 3745-89-03 of the Administrative Code, within the time frame specified by the director. The corrections shall be reported to the director in writing, satisfactorily addressing each individual deviation.

- (B) The laboratory may be required to issue notification to its customers, within a specified period of time, explaining deviations noted in the survey. The laboratory may also be required to notify affected customers of unacceptable data, which may include, but not be limited to, a clear, accurate and easily understood explanation of why the results were unacceptable.
  
- (C) In addition to paragraph (A) of this rule, laboratories holding a valid and unexpired laboratory certification for gross alpha particle activity, radium-226 and radium-228 radioactivity, gross beta particle activity, strontium-89 and strontium-90 radioactivity, tritium, or radioactivity from photon emitters (excluding iodine-131), shall participate in all quality control samples and proficiency test samples. Satisfactory performance in appropriate analyses as described in rule 3745-89-03 of the Administrative Code shall be obtained in at least one proficiency test study per year.
  
- (D) Failure to maintain certification in accordance with this rule may result in suspension or revocation of a laboratory's certification in accordance with rule 3745-89-06 of the Administrative Code.

Effective:

Five Year Review (FYR) Dates: 05/04/2020

---

Certification

---

Date

Promulgated Under: 119.03  
Statutory Authority: 6109.03, 6109.04  
Rule Amplifies: 6109.03, 6109.04, 3745.50  
Prior Effective Dates: 11/26/80, 04/09/90, 04/01/99, 06/18/04, 01/08/10,  
05/04/15